

Manual Supplement

Manual Title:	1760 Reference	Supplement Issue:	1
Print Date:	July 2006	Issue Date:	2/14
Revision/Date:	2, 7/10	Page Count:	2

This supplement contains information necessary to ensure the accuracy of the above manual. This manual is distributed as an electronic manual on the following CD-ROM:

CD Title:	1760
CD Rev. & Date:	5, 7/10
CD PN:	2583518

Change #1


On page 1-2, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
---	--

On page 5-7, replace the figure with:

Nominal and Limit Values

Vn - fn - IL	D - S - I	EN50160	EN50160 hxx	Rapid Voltage Changes
--------------	-----------	---------	-------------	-----------------------



Statistics Evaluation

Limit settings for statistical evaluation as per EN50160. Slow voltage variations, frequency, flicker, maximum number of events.

Slow Voltage Variations

Values within tolerance during 95% 99.0% of the measure time.

95% Limit	pos. <input type="text" value="110.00"/> %	neg. <input type="text" value="90.00"/> %
100% Limit	pos. <input type="text" value="110.00"/> %	neg. <input type="text" value="85.00"/> %

Fast Voltage Variations

95% Limit %

Mains Frequency

Values within tolerance during 95% 99.5% of the measure time.

95% Limit	pos. <input type="text" value="101.00"/> %	neg. <input type="text" value="99.00"/> %
100% Limit	pos. <input type="text" value="104.00"/> %	neg. <input type="text" value="94.00"/> %

Long-term Flicker (Plt)	<input type="text" value="1.00"/>
Max. number of events	<input type="text" value="100"/>
Unbalance	<input type="text" value="2.00"/> %

On page 6-23, add the following under the Note:

See below for a Dis/Dip table showing the distribution statistics of voltage dips depending on the length of the dip event according to EN50160:

EN50160 - Event Statistics					
Voltage Swells 3-ph:					
Swell Voltage [%Vnom]	Duration [ms]				
	½cyc <= t <= 500	500 < t <= 5000	5000 < t <= 60000		
u >= 120%	6	3	-		
120% > u >= 110%	4	-	-		
Voltage Dips 3-ph:					
Residual Voltage [%Vnom]	Duration [ms]				
	½cyc <= t <= 200	200 < t <= 500	500 < t <= 1000	1000 < t <= 5000	5000 < t <= 60000
90% > u >= 80%	-	3	-	-	-
80% > u >= 70%	3	-	-	-	-
70% > u >= 40%	-	-	-	-	-
40% > u >= 5%	-	-	-	-	-
5% > u >= 0%	-	-	-	-	1