

Manual Supplement

Manual Title:	1760 Getting Started	Supplement Issue:	2
Part Number:	2713103	Issue Date:	2/14
Print Date:	June 2006	Page Count:	2
Revision/Date:	2, 7/10		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 6, replace Figure 2 with:





Figure 2. Instrument Label

Change #2

On page 1, under **Safety Instructions**, replace the first sentence with:

The design and manufacture of the device conform to the latest state of technology and the safety standards laid down in IEC 61010-1 and IEC 61010-2-30.

On page 2, add the following to the Symbols table:

	Conforms to relevant South Korean EMC Standards.
---	--

On page 8, replace the 7th bullet with:

1 crossover Ethernet cable for direct-PC connection (RJ45 crossover cable, do not use the gray plug).

On page 14, under ***Communicating with the Recorder***, replace the second paragraph with:

Supply the Recorder with mains power, switch it on, connect a voltage sensor to channel 1 and connect the network interface to your PC using the Ethernet cable for direct connection (RJ45 crossover cable, do not use the gray plug).