

Manual Supplement

| | | | |
|----------------|----------------------|-------------------|----------|
| Manual Title: | 1760 Getting Started | Supplement Issue: | 2 |
| Part Number: | 2713103 | Issue Date: | 2/14 |
| Print Date: | June 2006 | Page Count: | 2 |
| Revision/Date: | 2, 7/10 | | |

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 6, replace Figure 2 with:





Figure 2. Instrument Label

Change #2

On page 1, under **Safety Instructions**, replace the first sentence with:

The design and manufacture of the device conform to the latest state of technology and the safety standards laid down in IEC 61010-1 and IEC 61010-2-30.

On page 2, add the following to the Symbols table:

| | |
|-----------------------------------------------------------------------------------|--------------------------------------------------|
|  | Conforms to relevant South Korean EMC Standards. |
|-----------------------------------------------------------------------------------|--------------------------------------------------|

On page 8, replace the 7th bullet with:

1 crossover Ethernet cable for direct-PC connection (RJ45 crossover cable, do not use the gray plug).

On page 14, under ***Communicating with the Recorder***, replace the second paragraph with:

Supply the Recorder with mains power, switch it on, connect a voltage sensor to channel 1 and connect the network interface to your PC using the Ethernet cable for direct connection (RJ45 crossover cable, do not use the gray plug).